

FRIEDMAN & BRUYA, INC.

ENVIRONMENTAL CHEMISTS

Analysis For Total Metals By EPA Method 200.8

Client ID:	M130742	Client:	Alaskan Copper Works
Date Received:	07/19/12	Project:	Metro Self Monitor M130742, F&BI 207260
Date Extracted:	07/24/12	Lab ID:	207260-01 x10
Date Analyzed:	07/24/12	Data File:	207260-01 x10.055
Matrix:	Water	Instrument:	ICPMS1
Units:	ug/L (ppb)	Operator:	AP

Internal Standard:	% Recovery:	Lower	Upper
Germanium	93	Limit:	Limit:
		60	125

Analyte:	Concentration ug/L (ppb)
Chromium	206
Nickel	198
Copper	435
Zinc	<50

FRIEDMAN & BRUYA, INC.

ENVIRONMENTAL CHEMISTS

Analysis For Total Metals By EPA Method 200.8

Client ID:	Method Blank	Client:	Alaskan Copper Works
Date Received:	Not Applicable	Project:	Metro Self Monitor M130742, F&BI 207260
Date Extracted:	07/24/12	Lab ID:	I2-488 mb
Date Analyzed:	07/24/12	Data File:	I2-488 mb.049
Matrix:	Water	Instrument:	ICPMS1
Units:	ug/L (ppb)	Operator:	AP

Internal Standard:	% Recovery:	Lower Limit:	Upper Limit:
Germanium	86	60	125

Analyte:	Concentration ug/L (ppb)
Chromium	<1
Nickel	<1
Copper	<1
Zinc	<5

FRIEDMAN & BRUYA, INC.

ENVIRONMENTAL CHEMISTS

Date of Report: 07/27/12

Date Received: 07/19/12

Project: Metro Self Monitor M130742, F&BI 207260

**QUALITY ASSURANCE RESULTS
FOR THE ANALYSIS OF WATER SAMPLES
FOR TOTAL METALS USING EPA METHOD 200.8**

Laboratory Code: 207300-01 (Matrix Spike)

Analyte	Reporting Units	Spike Level	Sample Result	Percent Recovery MS	Percent Recovery MSD	Acceptance Criteria	RPD (Limit 20)
Chromium	ug/L (ppb)	20	<1	104	103	71-130	1
Nickel	ug/L (ppb)	20	1.15	103	101	71-120	2
Copper	ug/L (ppb)	20	1.90	101	100	52-134	1
Zinc	ug/L (ppb)	50	6.70	105	103	51-142	2

Laboratory Code: Laboratory Control Sample

Analyte	Reporting Units	Spike Level	Percent Recovery LCS	Acceptance Criteria
Chromium	ug/L (ppb)	20	100	80-119
Nickel	ug/L (ppb)	20	102	83-119
Copper	ug/L (ppb)	20	102	81-120
Zinc	ug/L (ppb)	50	102	82-120

FRIEDMAN & BRUYA, INC.

ENVIRONMENTAL CHEMISTS

Data Qualifiers & Definitions

a - The analyte was detected at a level less than five times the reporting limit. The RPD results may not provide reliable information on the variability of the analysis.

A1 - More than one compound of similar molecule structure was identified with equal probability.

b - The analyte was spiked at a level that was less than five times that present in the sample. Matrix spike recoveries may not be meaningful.

ca - The calibration results for this range fell outside of acceptance criteria. The value reported is an estimate.

c - The presence of the analyte indicated may be due to carryover from previous sample injections.

d - The sample was diluted. Detection limits may be raised due to dilution.

ds - The sample was diluted. Detection limits are raised due to dilution and surrogate recoveries may not be meaningful.

dv - Insufficient sample was available to achieve normal reporting limits and limits are raised accordingly.

fb - Analyte present in the blank and the sample.

fc - The compound is a common laboratory and field contaminant.

hr - The sample and duplicate were reextracted and reanalyzed. RPD results were still outside of control limits. The variability is attributed to sample inhomogeneity.

ht - Analysis performed outside the method or client-specified holding time requirement.

ip - Recovery fell outside of normal control limits. Compounds in the sample matrix interfered with the quantitation of the analyte.

j - The result is below normal reporting limits. The value reported is an estimate.

J - The internal standard associated with the analyte is out of control limits. The reported concentration is an estimate.

jl - The analyte result in the laboratory control sample is out of control limits. The reported concentration should be considered an estimate.

jr - The rpd result in laboratory control sample associated with the analyte is out of control limits. The reported concentration should be considered an estimate.

js - The surrogate associated with the analyte is out of control limits. The reported concentration should be considered an estimate.

lc - The presence of the compound indicated is likely due to laboratory contamination.

L - The reported concentration was generated from a library search.

nm - The analyte was not detected in one or more of the duplicate analyses. Therefore, calculation of the RPD is not applicable.

pc - The sample was received in a container not approved by the method. The value reported should be considered an estimate.

pr - The sample was received with incorrect preservation. The value reported should be considered an estimate.

ve - Estimated concentration calculated for an analyte response above the valid instrument calibration range. A dilution is required to obtain an accurate quantification of the analyte.

vo - The value reported fell outside the control limits established for this analyte.

x - The sample chromatographic pattern does not resemble the fuel standard used for quantitation.

207.260

Send Report To: GEORGE THOMPSON
Company: ALASKA COPPER WORKS
Address: 628 S. HANFORD ST
City, State, ZIP: SEATTLE WA 98134
Phone # 206-571-6033 Fax # 206-382-4309

SAMPLE CHAIN OF CUSTODY ME 07-19-12

AIS

SAMPLE # (signature)	
PROJECT NAME NO.	PO #
METRO SELF MONITOR	M/30742
REMARKS	

Page # _____ of _____

TURNAROUND TIME

☐ Standard (2 Weeks)

☒ **EXPRESS 4 day**

Rush charges authorized by: _____

SAMPLE DISPOSAL

☐ Dispose after 30 days



☐ Return samples

☐ Will call with instructions

[illegible]

Friedman & Bruya, Inc.
3012 16th Avenue West
Seattle, WA 98119-2029
Ph. (206) 295-8282
Fax (206) 293-5044

FORMS\CGC\CGC.DOC

SIGNATURE	PRINT NAME	COMPANY	DATE	TIME
Relinquished by: 	Gerald T. Thompson	ACW	7/19/12	1:00pm
Received by: 	Nhan Phan	FeBT	7/19/12	1:26
Relinquished by:				
Received by:				

Samples received at 24 °C

FRIEDMAN & BRUYA, INC.

ENVIRONMENTAL CHEMISTS

James E. Bruya, Ph.D.
Charlene Morrow, M.S.
Yelena Aravkina, M.S.
Bradley T. Benson, B.S.
Kurt Johnson, B.S.

3012 16th Avenue West
Seattle, WA 98119-2029
TEL: (206) 285-8282
FAX: (206) 283-5044
e-mail: fbi@isomedia.com

July 27, 2012

Gerald Thompson, Project Manager
Alaskan Copper Works
628 South Hanford
Seattle, WA 98134

Dear Mr. Thompson:

Included are the results from the testing of material submitted on July 19, 2012 from the Metro Self Monitor M130742, F&BI 207260 project. There are 4 pages included in this report. Any samples that may remain are currently scheduled for disposal in 30 days. If you would like us to return your samples or arrange for long term storage at our offices, please contact us as soon as possible.

We appreciate this opportunity to be of service to you and hope you will call if you have any questions.

Sincerely,

FRIEDMAN & BRUYA, INC.



Michael Erdahl
Project Manager

Enclosures
ACU0727R.DOC